



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

09/631,000

08/02/2000

Michael L. Blomquist

9015.135US01

8019

23552 7590 06/22/2009  
MERCHANT & GOULD PC  
P.O. BOX 2903  
MINNEAPOLIS, MN 55402-0903

EXAMINER

KOPPIKAR, VIVEK D

ART UNIT

PAPER NUMBER

3686

MAIL DATE

DELIVERY MODE

06/22/2009

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 09/631,000	<b>Applicant(s)</b> BLOMQUIST, MICHAEL L.	
	<b>Examiner</b> VIVEK D. KOPPIKAR	<b>Art Unit</b> 3686	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 26 May 2009.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-19, 23-24 and 26-32 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-19, 23-24 and 26-32 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

## **DETAILED ACTION**

### ***Status of the Application***

1. Claims 1-19, 23-24 and 26-32 have been examined in this application. This communication is a Final Office Action in response to the "Remarks" section filed on May 26, 2009.

### ***Claim Rejections - 35 USC § 103***

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 1-3, 6-19 and 26-32 are rejected under 35 U.S.C. 103(a) as being unpatentable over US Patent Number 5,788,669 to Peterson in view of US Patent Number 5,713,856 to Eggers in view of in view of Official Notice and in even further view of US Patent Number 5,658,250 to Blomquist.

(A) As per claim 1, Peterson discloses a method for creating a library of pump data on a computer having a database, the pump data being organized into sets of program data, each set of program data being available for batch downloading to a medical pump and including data items for controlling operation of the medical pump, the method comprising:

the plurality of data items forming a set of program data, (col. 4, lines 10-18 and lines 36-53).

Peterson teaches parameters (Col. 1, Ln. 20-21). In Peterson the memory is within the pump (Figure 1 and Col. 3, Ln. 50-65).

Art Unit: 3686

Peterson does not explicitly disclose at least some of the data items establishing parameters for controlling operation of a medical pump entering a plurality of data items into a database on the computer.

However, Eggers discloses at least some of the data items establishing parameters for controlling operation of a medical pump (i.e. drug libraries customized for each user ...) (col. 10, line 62 - col. 11, line 45) entering a plurality of data being patient-specific data items into a database on the computer, and assigning at least one data key to the set of program data, the data key identifying the set of program data (i.e. drug libraries customized for each user ...)(col. 10, line 62 - col. 11, line 45). Eggers also teaches the following: batch-downloading the plurality of data items into the memory within the pump and controlling operation of the pump based on one or more data items (Eggers: Col. 10, Ln. 62-Col. 11, Ln. 7). It would have been obvious to one of ordinary skill in the art at the time of Applicant's invention to include at least some of the data items establishing parameters for controlling operation of a medical pump, entering a plurality of data items into a database on the computer as disclosed by Eggers within the Peterson system for the motivation of downloading complicated drug delivery profiles to the system (col. 2, lines 3-10 and col. 11, lines 14-20).

Peterson and Eggers do not explicitly disclose assigning at least one data key to the set of program data, the data key identifying the set of program data. However, the Examiner takes official notice that it was well known in the database arts to assign identifiers to data sets. The purpose of using identifiers was to locate the particular data that is to be utilized by a user or program. It would have been obvious to one of ordinary skill in the art at the time of Applicants

Art Unit: 3686

invention to include assigning at least one data key to the set of program data, the data key identifying the set of program data within Peterson and Eggers for the motivation stated above.

The combined teachings of Peterson in view of Eggers in view of Official Notice do not teach or suggest individualized, patient-specific data items, however, this feature is taught in Blomquist (Col. 11, Ln. 64-Col. 12, Ln. 10). At the time of the invention, it would have been obvious for one of ordinary skill in the art to have modified the combined teachings of Peterson in view of Eggers in view of Official Notice with these aforementioned teachings from Blomquist with the motivation of having a means of allowing a user to download applications into a pump's flash memory without having to enter the information through a keyboard, as recited in Blomquist (Col. 12, Ln. 4-11).

(B) As to claim 2, Peterson does not explicitly disclose the method of claim 1 wherein the acts of:

entering a plurality of data items into a database includes entering the plurality of data items into a program data record in the database.

However, Eggers discloses entering a plurality of data items into a database includes entering the plurality of data items into a program data record in the database (i.e. drug library (Col. 2, Ln. 3-10 and Col. 11, Ln. 14-26). It would have been obvious to one of ordinary skill in the art at the time of Applicant's invention to include entering a plurality of data items into a database includes entering the plurality of data items into a program data record in the database as disclosed by Eggers within the Peterson system for the motivation of downloading complicated drug delivery profiles to the system (col. 2, lines 3-10 and col. 11, lines 14-2). Peterson and Eggers do not explicitly disclose

Art Unit: 3686

assigning at least one data key to the set of program data includes entering the data key into a data key record and linking the data key record to the program data record.

However, the Examiner takes official notice that it was well known in the database arts to assign identifiers to data sets and linking data key records to application programs.. The purpose of using identifiers was to locate the particular data that is to be utilized by a user or program. It would have been obvious to one of ordinary skill in the art at the time of applicant's invention to include assigning at least one data key to the set of program data includes entering the data key into a data key record and linking the data key record to the program data record within Peterson and Eggers for the motivation stated above.

(C) As to claim 3, Peterson does not explicitly disclose the method of claim 2 wherein further including entering an identification code selected from the group consisting essentially of a patient I.D., a therapy I.D., and a fluid I.D., wherein the patient I.D. is a code identifying a patient, the therapy I.D. is a code identifying a therapy administered using a medical pump, and the fluid I.D. is a code identifying a fluid that is administered using a medical pump.

However, Eggers discloses further including entering an identification code selected from the group consisting essentially of a patient I.D., a therapy I.D., and a fluid I.D., wherein the patient I.D. is a code identifying a patient, the therapy I.D. is a code identifying a therapy administered using a medical pump, and the fluid I.D. is a code identifying a fluid that is administered using a medical pump (col. 10, line 62 - col. 11, line 7). It would have been obvious to one of ordinary skill in the art at the time of Applicant's invention to include entering an identification code selected from the group consisting essentially of a patient I.D., a therapy I.D., and a fluid I.D., wherein the patient I.D. is a code identifying a patient, the therapy I.D. is a code

Art Unit: 3686

identifying a therapy administered using a medical pump, and the fluid I.D. is a code identifying a fluid that is administered using a medical pump as disclosed by Eggers within the Peterson system for the motivation of downloading complicated drug delivery profiles to the system (col. 2, lines 3-10 and col. 11, lines 14-20).

(D) As to claim 6, Peterson in view of Eggers in view of Blomquist teach a computer storage medium containing a library of pump data, the computer storage medium being created by the method set forth in claim 1.

(E) Claims 4 and 5 are rejected under 35 U.S.C. 103(a) as being unpatentable over Peterson in view Eggers in view of Blomquist in view of Official Notice as applied to claim 3 above, and further in view of “Acute Health Solutions” DoseWatch to use Multum's MediSource (hereinafter Medisource).

As to claim 4, Peterson does not explicitly disclose the method of claim 3 wherein the computer is in data communication with a scanner, the method further comprising.

scanning a bar code with the scanner; and

entering the bar code into the computer, wherein the act of assigning at least one data key to the

set of program data includes assigning the bar code to the set of program data.

However, MediSource discloses wherein the computer is in data communication with a scanner, the method further comprising.

scanning a bar code with the scanner; and entering the bar code into the computer, wherein the act of assigning at least one data key to the set of program data includes assigning the bar code to the set of program data (see abstract and

Art Unit: 3686

page 2, paragraph 3). It would have been obvious to one of ordinary skill in the art at the time of Applicant's invention to include wherein the computer is in data communication with a scanner, the method further comprising: scanning a bar code with the scanner; and entering the bar code into the computer, wherein the act of assigning at least one data key to the set of program data includes assigning the bar code to the set of program data as disclosed by MediSource with the combined system of Peterson, Eggers, Blomquist and Official Notice for the motivation to insure association of the drug and concentration with a pump rate and an infusion amount (Abstract).

As to claim 5, Peterson does not explicitly disclose the method of claim 3 wherein the computer is in data communication with a medical pump, the method further comprising uploading a set of program data items from the pump.

However, MediSource discloses wherein the computer is in data communication with a medical pump, the method further comprising uploading a set of program data items from the pump (Abstract and Page 2). It would have been obvious to one of ordinary skill in the art at the time of the invention to include a computer which is in data communication with a medical pump, the method further comprising uploading a set of program data items from the pump as disclosed by MediSource within the Peterson system for the motivation of insuring association of the drug and concentration with a pump rate and an infusion amount (Abstract).

(F) As per claims 7-19 and 23-24, these claims are substantially similar to claims 1-6 and are rejected on the same basis.

(G) Claims 26-31 are rejected under 35 U.S.C. 103(a) as being unpatentable over Peterson in view of Eggers in view of Blomquist.



Art Unit: 3686

As per claim 26, Peterson teaches a pump for infusing fluid into a patient (Peterson: Figure 1 and Col. 3, Ln. 50-65), the pump comprising:

- a housing (Figure 1 and Col. 3, Ln. 50-65);
- a pump mechanism positioned within the housing (Figure 1 and Col. 3, Ln. 50-65);
- memory positioned within the housing and configured to store a plurality of data items being patient-specific data items forming a set of program data, at least some of the data items establishing patient-specific parameters for controlling operation of a medical pump (Figure 1; Col. 1, Ln. 1-21 and Col. 3, Ln. 50-65)

and multiple program modules (Col. 3, Ln. 56-65 and Col. 4, Ln. 54-57).

Peterson does not explicitly disclose at least some of the data items establishing parameters for controlling operation of a medical PUDP entering a plurality of data items into a database on the computer.

However, Eggers discloses at least some of the data items establishing parameters for controlling operation of a medical pump (i.e. drug libraries customized for each user ...) (col. 10, line 62 - col. 11, line 45) entering a plurality of data items into a database on the computer, and assigning at least one data key to the set of program data, the data key identifying the set of program data (i.e. drug libraries customized for each user ...)(col. 10, line 62 - col. 11, line 45).

The combined teachings of Peterson in view of Eggers do not teach or suggest individualized, patient-specific data items, however, this feature is taught in Blomquist (Col. 11, Ln. 64-Col. 12, Ln. 10). At the time of the invention, it would have been obvious for one of ordinary skill in the art to have modified the combined teachings of Peterson in view of Eggers in view of Official Notice with these aforementioned teachings from Blomquist with the

Art Unit: 3686

motivation of having a means of allowing a user to download applications into a pump's flash memory without having to enter the information through a keyboard, as recited in Blomquist (Col. 12, Ln. 4-11).

(H) As per claim 27, in Peterson the program data includes the delivery rate for the patient (Peterson: Col. 4, Ln. 47-53).

(I) As per claim 28, in Peterson the program data identifies a therapy name (Peterson: Col. 4, Ln. 47-53).

(J) As per claims 29-31, are these claims are substantially similar to claims 1-20, 23-24 and 26-28, above, are rejected on the same basis.

(K) As per claim 32, the combined teachings of Peterson in view of Eggers teaches that the plurality of data items includes at least one data item selected from the group consisting of data items related to delivery schedules, medication doses, and boluses (Eggers: Col. 10, Ln. 62-66).

The motivation for making this modification to Peterson is the same as that set forth above, in the rejection of Claim 26.

### ***Response to Arguments***

4. Applicant's arguments filed May 26, 2009 have been fully considered but they are not persuasive. Applicants arguments will be addressed in sequential order as they were presented in the "Remarks" section filed on May 26, 2009.

(1) The applicants argue that the Office does not include any analysis or discussion about why it would have been obvious to modify the other cited references in view of Blomquist to batch download individualized, patient-specific data. However, the Office does in fact

Art Unit: 3686

provide the following motivation statement as being a reason that would motivate one of ordinary skill in the art to modify the cited references with the teachings from Blomquist:

“The combined teachings of Peterson in view of Eggers in view of Official Notice do not teach or suggest individualized, patient-specific data items, however, this feature is taught in Blomquist (Col. 11, Ln. 64-Col. 12, Ln. 10). At the time of the invention, it would have been obvious for one of ordinary skill in the art to have modified the combined teachings of Peterson in view of Eggers in view of Official Notice with these aforementioned teachings from Blomquist with the motivation of having a means of allowing a user to download applications into a pump's flash memory without having to enter the information through a keyboard, as recited in Blomquist (Col. 12, Ln. 4-11).”

(2) Applicants argue that the Eggers and the Blomquist reference teach away from a claimed combination of batch downloading and individualized, patient-specific data items. However, the Office would like to point out that the Blomquist reference clearly states the limitation of downloading individualized, patient-specific data items (Blomquist: Col. 11, Ln. 64-Col. 12, Ln. 10). This cited portion of Blomquist teaches batch downloading individualized, patient specific parameters and it states that one of the advantages of doing so is so that numerous, patient-specific inputs would not have to be individually entered or entered using the keyboard.

Applicants also state that in Col. 10, lines 15-17 of Blomquist there is a discussion of using prompts for entering a “particular setting.” However, the Office would like to point out that despite this disclosure in Blomquist, the Blomquist reference goes on to disclose that

Art Unit: 3686

individualized, patient-specific can be batch-downloaded into a memory (Blomquist: Col. 11, Ln. 64-Col. 12, Ln. 10).

As a further note the Office would like to point out that Applicants make the following statement in their response (Page 3, 1<sup>st</sup> Full Paragraph) “[a]lthough the claims may not be so limited, Blomquist also does not teach that any patient specific data is batch downloaded.” Therefore it seems that the applicants are admitting that the claims of the Blomquist reference disclose the embodiment of batch downloading individualized, patient-specific data.

(3) Applicants argue that the Eggers reference explicitly distinguished between loading libraries and data such as RATE and VTBI into a pump and that in Eggers a drug library is loaded in a pump with a PCMIA interface or similar data communication interface, while settings such as RATE and VTBI are entered directly and individually into the pump as discussed above. The applicants then go on to argue that both references teach individually entering data such as RATE, not batch-downloading such data items.

To respond to this argument the Office would like to point out that the Blomquist reference is relied on to show that the feature of batch-downloading individualized, patient-specific data is known in the prior art and the Eggers reference is not relied upon for this purpose.

Applicants also argue that both Eggers and Blomquist teach individually entering such data such as RATE, not batch-downloading such data items. However, the Office would like to point out that the claims do not explicitly claim the step of batch-downloading RATE. The claims implicitly recite the step of "batch-downloading individualized, patient-specific data items" and as noted above this step is taught in Blomquist (Col. 11, Ln. 64-Col. 12, Ln. 10).

***Conclusion***

5. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

6. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Vivek Koppikar, whose telephone number is (571) 272-5109. The examiner can normally be reached from Monday to Friday between 8 AM and 4:30 PM.

If any attempt to reach the examiner by telephone is unsuccessful, the examiner's supervisor, Jerry O'Connor, can be reached at (571) 272-6787. The fax telephone numbers for this group are either (571) 273-8300 or (703) 872-9326 (for official communications including After Final communications labeled "Box AF").

Another resource that is available to applicants is the Patent Application Information Retrieval (PAIR). Information regarding the status of an application can be obtained from the (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAX. Status information for unpublished applications is available

Art Unit: 3686

through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, please feel free to contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Sincerely,

/Vivek D Koppikar/

Primary Examiner, Art Unit 3686

6/19/2009